

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1666

Name: RITTER2  
Sterilizer ID: V2313228  
11/02/2022 07:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.8 F 0.0  
2:00 74.1 F 0.0  
4:00 78.0 F 0.0  
6:00 103.2 F 0.1  
8:00 140.0 F 0.3  
10:00 172.9 F 0.4  
12:00 197.0 F 2.7  
14:00 211.4 F 6.3  
16:00 223.5 F 10.0  
18:00 231.0 F 9.8  
20:00 235.3 F 10.1  
22:00 237.1 F 9.3  
24:00 245.6 F 13.5  
26:00 244.4 F 12.5  
28:00 256.5 F 19.0  
30:00 267.4 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.4 F 29.0  
2:00 272.3 F 29.3  
3:00 271.9 F 28.6  
4:00 272.7 F 29.3

Min 271.6 F 28.5  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.7 F 0.0  
5:00 222.2 F 0.0  
10:00 226.1 F 0.0  
15:00 225.4 F 0.0  
20:00 229.8 F 0.0  
25:00 231.8 F 0.0  
30:00 229.1 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:25  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:09:18  
CYCLE COMPLETE

11/02/2022 08:52 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_