

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2694

Name: RITTER2  
Sterilizer ID: V2313228  
07/25/2023 07:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.7 F 0.0  
2:00 73.8 F 0.0  
4:00 80.5 F 0.0  
6:00 102.0 F 0.0  
8:00 154.2 F 0.3  
10:00 185.6 F 1.9  
12:00 211.7 F 5.6  
14:00 215.7 F 9.6  
16:00 230.6 F 11.8  
18:00 233.1 F 9.3  
20:00 238.0 F 10.4  
22:00 242.7 F 12.2  
24:00 240.8 F 10.9  
26:00 254.2 F 17.7  
28:00 265.6 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.6  
1:00 272.4 F 29.4  
2:00 271.4 F 28.5  
3:00 272.5 F 29.2  
4:00 271.5 F 28.7

Min 271.2 F 28.4  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.6 F 0.0  
5:00 221.2 F 0.0  
10:00 179.1 F 0.0  
15:00 176.9 F 0.0  
20:00 174.6 F 0.0  
25:00 174.2 F 0.0  
30:00 175.1 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 29:36  
STERILIZING: 4:00  
VENTING: 2:56  
DRYING: 30:00  
TOTAL CYCLE: 01:07:21  
CYCLE COMPLETE

07/25/2023 09:05 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_