
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002022
11:15 12/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:36
STERILIZING 3:42
136.4 C 225kPa 3:42
136.2 C 224kPa 4:12
136.4 C 226kPa 4:42
136.2 C 224kPa 5:12
136.4 C 226kPa 5:42
136.4 C 226kPa 6:12
136.4 C 226kPa 6:42
136.4 C 226kPa 7:12
136.4 C 226kPa 7:12
VENTING 7:12
AIR DRYING 8:04
CYCLE COMPLETE 21:07

Digital Signature #
B8346349FF9B6FBB
