
Midmark M11 - v1.0.5
Total Cycles: 244

Name: RITTER2
Sterilizer ID: V2313228
07/26/2021 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.7 F 0.0
2:00 158.7 F 2.9
4:00 182.9 F 6.9
6:00 206.3 F 12.0
8:00 217.8 F 6.6
10:00 231.7 F 11.8
12:00 234.0 F 9.3
14:00 237.8 F 9.8
16:00 248.3 F 14.7
18:00 260.3 F 21.5
20:00 271.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.4 F 30.0
2:00 272.4 F 29.9
3:00 272.0 F 29.4
4:00 272.2 F 29.6

Min 270.9 F 28.7
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.0 F 0.0
5:00 198.7 F 0.0
10:00 201.4 F 0.0
15:00 200.6 F 0.0
20:00 198.1 F 0.0
25:00 178.8 F 0.0
30:00 171.2 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 20:41
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:09
CYCLE COMPLETE

07/26/2021 10:21 AM
ACCEPTED: X
REJECTED:
Operator: U01