
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000466
14:50 16/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:42
STERILIZING 3:49
136.3 C 223kPa 3:49
136.3 C 225kPa 4:19
136.1 C 220kPa 4:49
136.4 C 225kPa 5:19
136.4 C 222kPa 5:49
136.4 C 223kPa 6:19
136.4 C 223kPa 6:49
136.1 C 223kPa 7:19
136.1 C 223kPa 7:19
VENTING 7:19
AIR DRYING 8:08
CYCLE COMPLETE 8:15

Digital Signature #
A8FF78A584EAF9EE
