
Midmark M11 - v1.0.0
Total Cycles: 1770

Name: RITTER 1
Sterilizer ID: V2178198
03/30/2022 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.0 F 0.0
2:00 188.5 F 3.6
4:00 206.2 F 8.0
6:00 225.5 F 10.6
8:00 232.9 F 9.2
10:00 237.1 F 9.9
12:00 243.4 F 12.4
14:00 241.9 F 11.1
16:00 255.0 F 17.8
18:00 266.6 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.4 F 28.8
2:00 271.5 F 28.1
3:00 272.4 F 29.0
4:00 271.9 F 28.5

Min 271.4 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 228.5 F 0.0
10:00 234.5 F 0.0
15:00 230.1 F 0.0
20:00 228.6 F 0.0
25:00 230.9 F 0.0
30:00 229.5 F 0.0

Drying Complete

FILLING: 1:55
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

03/30/2022 01:58 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____