
Midmark M11 - v1.0.5
Total Cycles: 3096

Name: RITTER2
Sterilizer ID: V2313228
11/07/2023 05:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.0 F 0.0
2:00 123.4 F 0.9
4:00 147.2 F 3.7
6:00 177.6 F 7.7
8:00 196.4 F 12.6
10:00 217.3 F 6.7
12:00 231.0 F 11.9
14:00 232.6 F 8.9
16:00 238.7 F 10.2
18:00 246.0 F 13.9
20:00 258.7 F 20.8
22:00 270.0 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.6 F 29.5
2:00 272.2 F 29.9
3:00 271.9 F 29.2
4:00 272.5 F 29.9

Min 271.4 F 29.1
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 202.7 F 0.0
10:00 197.9 F 0.0
15:00 197.7 F 0.0
20:00 194.3 F 0.0
25:00 193.1 F 0.0
30:00 191.8 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 22:51
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:01:12
CYCLE COMPLETE

11/07/2023 06:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____