
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002094
13:02 24/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:14
STERILIZING 4:15
136.3 C 223kPa 4:15
136.4 C 225kPa 4:45
136.4 C 225kPa 5:15
136.4 C 224kPa 5:45
136.1 C 223kPa 6:15
136.3 C 225kPa 6:45
136.4 C 225kPa 7:15
136.1 C 222kPa 7:45
136.1 C 222kPa 7:45
VENTING 7:45
AIR DRYING 8:41
CYCLE COMPLETE 11:07

Digital Signature #
D0F02C36688BC55D
