
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008103
11:02 27/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:25
STERILIZING 3:28
136.4 C 222kPa 3:28
136.3 C 222kPa 3:58
136.2 C 223kPa 4:28
136.4 C 222kPa 4:58
136.1 C 220kPa 5:28
136.4 C 223kPa 5:58
135.8 C 220kPa 6:28
136.4 C 221kPa 6:58
136.4 C 221kPa 6:58
VENTING 6:58
AIR DRYING 7:49
CYCLE COMPLETE 10:10

Digital Signature #
C92812C9FBD93D04
