
Midmark M11 - v1.0.0
Total Cycles: 2224

Name: RITTER 1
Sterilizer ID: V2178198
08/15/2022 02:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.3 F 0.0
2:00 155.4 F 3.1
4:00 172.9 F 7.1
6:00 194.2 F 11.6
8:00 214.5 F 5.3
10:00 227.3 F 10.3
12:00 231.6 F 7.5
14:00 240.6 F 11.9
16:00 238.3 F 9.8
18:00 251.5 F 16.2
20:00 263.0 F 22.9
22:00 272.2 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.4 F 28.9
2:00 271.6 F 28.7
3:00 272.2 F 28.9
4:00 271.5 F 28.7

Min 271.4 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 228.6 F 0.0
10:00 232.7 F 0.0
15:00 230.1 F 0.0
20:00 227.1 F 0.0
25:00 222.8 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 22:12
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:26
CYCLE COMPLETE

08/15/2022 03:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____