

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1721

Name: RITTER2  
Sterilizer ID: V2313228  
11/14/2022 05:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 206.1 F 0.0  
2:00 202.9 F 3.4  
4:00 218.3 F 9.3  
6:00 227.9 F 8.2  
8:00 237.7 F 11.9  
10:00 237.9 F 10.4  
12:00 238.2 F 9.6  
14:00 252.5 F 16.9  
16:00 264.8 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.3 F 29.1  
2:00 272.0 F 29.5  
3:00 272.0 F 28.9  
4:00 272.2 F 29.6  
  
Min 271.3 F 28.7  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.6 F 0.0  
5:00 226.2 F 0.0  
10:00 233.4 F 0.0  
15:00 228.9 F 0.0  
20:00 229.4 F 0.0  
25:00 228.1 F 0.0  
30:00 226.7 F 0.0  
  
Drying Complete

FILLING: 2:31  
HEATING: 17:48  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:30  
CYCLE COMPLETE

11/14/2022 06:37 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_