
Midmark M11 - v1.0.0
Total Cycles: 5210

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2024 11:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.0 F 0.0
2:00 149.3 F 0.0
4:00 162.7 F 0.1
6:00 185.5 F 0.4
8:00 206.5 F 3.4
10:00 216.3 F 7.2
12:00 228.6 F 11.3
14:00 228.7 F 6.6
16:00 240.8 F 12.1
18:00 238.9 F 9.9
20:00 238.0 F 8.8
22:00 250.6 F 15.2
24:00 262.0 F 21.6
26:00 271.7 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.1 F 28.4
2:00 272.5 F 28.6
3:00 272.1 F 28.2
4:00 271.7 F 27.9

Min 271.4 F 27.8
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.5 F 0.0
5:00 213.3 F 0.0
10:00 233.3 F 0.0
15:00 216.1 F 0.0
20:00 211.4 F 0.0
25:00 214.9 F 0.0
30:00 214.4 F 0.0
Drying Complete

FILLING: 0:42
HEATING: 26:44
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:05:15
CYCLE COMPLETE

12/20/2024 12:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____