

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 007755  
14:14 15/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:08  
PRESSURIZING 2:12  
STERILIZING 3:08  
136.0 C 220kPa 3:08  
136.4 C 225kPa 3:38  
136.3 C 223kPa 4:08  
136.1 C 221kPa 4:38  
136.4 C 223kPa 5:08  
136.4 C 224kPa 5:38  
136.4 C 224kPa 6:08  
136.4 C 224kPa 6:38  
136.4 C 224kPa 6:38  
VENTING 6:38  
AIR DRYING 7:30  
CYCLE COMPLETE 15:54

-----  
Digital Signature #  
EDBF00D8D1A15782  
-----