
Midmark M11 - v1.0.5
Total Cycles: 117

Name: RITTER2
Sterilizer ID: V2313228
06/18/2021 10:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.7 F 0.0
2:00 134.6 F 1.3
4:00 141.9 F 3.6
6:00 166.4 F 6.8
8:00 188.8 F 10.4
10:00 199.7 F 6.7
12:00 213.4 F 6.9
14:00 225.9 F 11.2
16:00 229.9 F 7.8
18:00 238.9 F 13.1
20:00 241.2 F 11.8
22:00 252.1 F 17.5
24:00 261.7 F 23.5
26:00 271.6 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.8
1:00 272.7 F 30.7
2:00 272.2 F 30.1
3:00 271.7 F 29.9
4:00 272.4 F 30.7

Min 271.6 F 29.7
Max 272.9 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.0 F 0.0
5:00 223.9 F 0.0
10:00 218.9 F 0.0
15:00 221.2 F 0.0
20:00 219.6 F 0.0
25:00 215.2 F 0.0
30:00 215.9 F 0.0

Drying Complete

FILLING: 1:02
HEATING: 26:40
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:05:12
CYCLE COMPLETE

06/18/2021 11:15 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____