
Midmark M11 - v1.0.0
Total Cycles: 1645

Name: RITTER 1
Sterilizer ID: V2178198
02/17/2022 07:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.5 F 0.0
2:00 95.4 F 0.0
4:00 117.1 F 0.0
6:00 157.8 F 1.1
8:00 176.9 F 2.5
10:00 190.6 F 4.9
12:00 214.4 F 8.0
14:00 223.0 F 11.5
16:00 219.7 F 4.4
18:00 231.0 F 8.8
20:00 241.5 F 13.0
22:00 236.7 F 9.5
24:00 238.1 F 9.6
26:00 245.5 F 13.0
28:00 257.4 F 19.4
30:00 268.3 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.0 F 28.7
2:00 272.1 F 29.1
3:00 271.6 F 28.4
4:00 272.2 F 29.1

Min 271.3 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.1 F 0.0
5:00 200.3 F 0.0
10:00 188.7 F 0.0
15:00 184.2 F 0.0
20:00 215.5 F 0.0
25:00 240.4 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 31:14
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:09:17
CYCLE COMPLETE

02/17/2022 09:17 AM
ACCEPTED: X
REJECTED:
Operator: U02