
Midmark M11 - v1.0.0
Total Cycles: 2241

Name: RITTER 1
Sterilizer ID: V2178198
08/24/2022 02:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.3 F 0.0
2:00 192.7 F 0.4
4:00 200.9 F 1.8
6:00 221.4 F 6.7
8:00 235.5 F 11.7
10:00 233.1 F 8.3
12:00 242.7 F 11.3
14:00 244.8 F 12.7
16:00 243.7 F 11.9
18:00 257.1 F 18.9
20:00 268.6 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.5 F 28.7
2:00 271.7 F 28.3
3:00 272.4 F 28.6
4:00 271.5 F 28.1

Min 271.5 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 231.9 F 0.0
10:00 226.6 F 0.0
15:00 229.6 F 0.0
20:00 227.3 F 0.0
25:00 228.0 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:10
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:30
CYCLE COMPLETE

08/24/2022 03:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____