
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002324
15:08 21/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:17
PRESSURIZING 2:27
STERILIZING 3:45
136.5 C 226kPa 3:45
136.0 C 222kPa 4:15
136.1 C 223kPa 4:45
136.2 C 224kPa 5:15
136.3 C 224kPa 5:45
136.4 C 224kPa 6:15
136.3 C 224kPa 6:45
136.4 C 225kPa 7:15
136.4 C 225kPa 7:15
VENTING 7:15
AIR DRYING 8:08
CYCLE COMPLETE 25:44

Digital Signature #
D9E04686AD03055B
