

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2223

Name: RITTER2  
Sterilizer ID: V2313228  
04/10/2023 12:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 215.3 F 0.0  
2:00 198.3 F 3.6  
4:00 207.5 F 8.9  
6:00 225.9 F 8.3  
8:00 235.0 F 11.0  
10:00 235.6 F 9.1  
12:00 238.7 F 9.6  
14:00 249.5 F 15.5  
16:00 262.0 F 22.7  
18:00 272.7 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.0  
1:00 272.6 F 29.4  
2:00 272.2 F 29.6  
3:00 272.3 F 29.2  
4:00 272.2 F 29.7  
  
Min 271.6 F 28.9  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.6 F 0.0  
5:00 231.1 F 0.0  
10:00 252.6 F 0.0  
15:00 231.7 F 0.0  
20:00 224.3 F 0.0  
25:00 221.6 F 0.0  
30:00 217.3 F 0.0  
Drying Complete

FILLING: 2:26  
HEATING: 18:15  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:57:53  
CYCLE COMPLETE

04/10/2023 01:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_