
Midmark M11 - v1.0.0
Total Cycles: 1867

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2022 05:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.9 F 0.0
2:00 184.0 F 4.0
4:00 202.8 F 8.7
6:00 223.4 F 7.3
8:00 231.8 F 9.4
10:00 237.9 F 9.0
12:00 243.2 F 12.6
14:00 242.0 F 11.4
16:00 255.6 F 18.4
18:00 267.6 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 28.9
2:00 271.5 F 28.6
3:00 272.4 F 28.9
4:00 271.3 F 28.4

Min 271.3 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 228.7 F 0.0
10:00 224.1 F 0.0
15:00 231.9 F 0.0
20:00 238.9 F 0.0
25:00 230.6 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 19:18
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

04/25/2022 06:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____