
Midmark M11 - v1.0.5
Total Cycles: 3026

Name: RITTER2
Sterilizer ID: V2313228
10/18/2023 12:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.5 F 0.0
2:00 180.0 F 3.9
4:00 199.7 F 9.3
6:00 217.3 F 6.8
8:00 228.7 F 10.1
10:00 232.9 F 8.1
12:00 240.9 F 11.1
14:00 247.0 F 14.2
16:00 260.2 F 21.5
18:00 272.0 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.7
2:00 271.7 F 28.9
3:00 272.5 F 29.8
4:00 272.4 F 29.4

Min 271.4 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 220.6 F 0.0
10:00 220.1 F 0.0
15:00 217.4 F 0.0
20:00 217.6 F 0.0
25:00 221.6 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 18:33
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:35
CYCLE COMPLETE

10/18/2023 01:01 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____