
Midmark M11 - v1.0.0
Total Cycles: 2815

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2023 03:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.1 F 0.0
2:00 185.7 F 3.4
4:00 203.2 F 8.4
6:00 220.8 F 9.8
8:00 230.5 F 9.6
10:00 234.6 F 8.5
12:00 245.2 F 13.5
14:00 244.8 F 12.5
16:00 258.0 F 19.6
18:00 269.6 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 28.9
2:00 271.3 F 28.2
3:00 272.6 F 29.0
4:00 271.9 F 28.5

Min 271.2 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 220.7 F 0.0
10:00 223.7 F 0.0
15:00 233.6 F 0.0
20:00 224.5 F 0.0
25:00 221.8 F 0.0
30:00 221.0 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 18:58
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

01/24/2023 04:28 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____