
Midmark M11 - v1.0.0
Total Cycles: 3192

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2023 08:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.1 F 0.0
2:00 189.0 F 1.0
4:00 196.6 F 4.3
6:00 208.5 F 9.6
8:00 223.2 F 7.2
10:00 235.0 F 10.2
12:00 235.0 F 7.8
14:00 243.6 F 12.2
16:00 244.7 F 12.8
18:00 257.7 F 19.7
20:00 269.1 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.6 F 29.3
2:00 271.9 F 28.7
3:00 271.5 F 28.8
4:00 272.6 F 29.3

Min 271.3 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 227.2 F 0.0
10:00 230.6 F 0.0
15:00 229.8 F 0.0
20:00 229.4 F 0.0
25:00 229.7 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 21:06
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

05/03/2023 09:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____