

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3340

Name: RITTER2  
Sterilizer ID: V2313228  
01/31/2024 12:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 108.9 F 0.0  
2:00 110.1 F 0.0  
4:00 127.8 F 1.2  
6:00 165.0 F 4.1  
8:00 180.7 F 7.8  
10:00 198.6 F 11.8  
12:00 220.3 F 5.3  
14:00 226.7 F 9.5  
16:00 238.7 F 10.8  
18:00 239.4 F 11.0  
20:00 236.4 F 9.2  
22:00 249.9 F 15.4  
24:00 261.0 F 21.7  
26:00 271.2 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.4 F 29.9  
2:00 271.8 F 29.8  
3:00 271.6 F 29.2  
4:00 272.9 F 29.6

Min 270.7 F 28.7  
Max 273.3 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.2 F 0.0  
5:00 235.0 F 0.0  
10:00 232.6 F 0.0  
15:00 212.2 F 0.0  
20:00 179.4 F 0.0  
25:00 176.6 F 0.0  
30:00 175.3 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 26:40  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:05:11  
CYCLE COMPLETE

01/31/2024 01:38 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_