

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3480

Name: RITTER2  
Sterilizer ID: V2313228  
03/27/2024 09:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.4 F 0.0  
2:00 129.7 F 0.2  
4:00 136.6 F 2.1  
6:00 161.6 F 5.3  
8:00 185.2 F 9.1  
10:00 202.8 F 13.3  
12:00 217.4 F 6.5  
14:00 228.9 F 10.9  
16:00 229.5 F 7.4  
18:00 240.9 F 12.7  
20:00 240.3 F 11.2  
22:00 252.0 F 17.0  
24:00 262.4 F 23.1  
26:00 272.5 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 272.8 F 29.9  
2:00 271.7 F 29.2  
3:00 271.6 F 29.6  
4:00 272.4 F 29.8

Min 271.2 F 29.1  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.3 F 0.0  
5:00 201.4 F 0.0  
10:00 195.1 F 0.0  
15:00 193.1 F 0.0  
20:00 190.4 F 0.0  
25:00 188.9 F 0.0  
30:00 188.1 F 0.0

Drying Complete

FILLING: 0:58  
HEATING: 26:28  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:05:01  
CYCLE COMPLETE

03/27/2024 10:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_