

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4240

Name: RITTER 1  
Sterilizer ID: V2178198  
01/29/2024 11:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.9 F 0.0  
2:00 151.3 F 1.8  
4:00 159.1 F 4.7  
6:00 181.3 F 8.4  
8:00 208.9 F 13.4  
10:00 222.4 F 6.4  
12:00 231.1 F 10.7  
14:00 231.2 F 7.2  
16:00 242.8 F 12.1  
18:00 239.4 F 10.0  
20:00 251.8 F 16.1  
22:00 262.7 F 22.3  
24:00 272.7 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.7 F 28.7  
2:00 272.6 F 29.1  
3:00 271.5 F 28.6  
4:00 272.7 F 28.8

Min 271.4 F 28.4  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.5 F 0.0  
5:00 237.3 F 0.0  
10:00 233.1 F 0.0  
15:00 223.4 F 0.0  
20:00 219.6 F 0.0  
25:00 220.9 F 0.0  
30:00 216.3 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 25:03  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:03:45  
CYCLE COMPLETE

01/29/2024 12:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_