
Midmark M11 - v1.0.5
Total Cycles: 2065

Name: RITTER2
Sterilizer ID: V2313228
02/16/2023 05:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.4 F 0.0
2:00 115.8 F 0.0
4:00 141.1 F 0.0
6:00 193.8 F 0.1
8:00 207.2 F 3.3
10:00 218.7 F 7.0
12:00 228.6 F 10.9
14:00 230.2 F 6.8
16:00 240.3 F 11.8
18:00 237.7 F 9.7
20:00 237.8 F 9.2
22:00 249.3 F 15.0
24:00 261.0 F 21.7
26:00 271.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.8
1:00 272.1 F 29.5
2:00 272.3 F 29.1
3:00 271.7 F 29.1
4:00 272.6 F 29.4

Min 271.5 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 210.1 F 0.0
10:00 209.5 F 0.0
15:00 221.0 F 0.0
20:00 235.0 F 0.0
25:00 235.8 F 0.0
30:00 225.7 F 0.0

Drying Complete

FILLING: 0:56
HEATING: 26:32
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:04:39
CYCLE COMPLETE

02/16/2023 06:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____