
Midmark M11 - v1.0.5
Total Cycles: 3914

Name: RITTER2
Sterilizer ID: V2313228
09/06/2024 03:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.4 F 0.0
2:00 136.4 F 0.3
4:00 152.1 F 2.8
6:00 186.0 F 6.7
8:00 207.5 F 11.0
10:00 217.9 F 5.3
12:00 232.3 F 10.0
14:00 233.7 F 9.3
16:00 242.2 F 12.4
18:00 240.6 F 11.2
20:00 253.2 F 17.4
22:00 263.9 F 23.8
24:00 272.0 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.5 F 29.5
2:00 271.9 F 29.6
3:00 272.1 F 29.4
4:00 272.6 F 29.7

Min 270.9 F 28.8
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 235.9 F 0.0
10:00 205.9 F 0.0
15:00 188.6 F 0.0
20:00 183.4 F 0.0
25:00 180.3 F 0.0
30:00 179.0 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 24:10
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:02:46
CYCLE COMPLETE

09/06/2024 04:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____