

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4411

Name: RITTER 1  
Sterilizer ID: V2178198  
03/27/2024 12:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.0 F 0.0  
2:00 162.5 F 3.0  
4:00 176.1 F 6.7  
6:00 193.9 F 10.8  
8:00 211.1 F 4.5  
10:00 226.8 F 8.9  
12:00 237.0 F 13.6  
14:00 237.6 F 10.4  
16:00 235.5 F 8.2  
18:00 248.7 F 14.6  
20:00 259.5 F 20.6  
22:00 269.5 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.5  
1:00 272.7 F 29.3  
2:00 272.2 F 28.8  
3:00 271.2 F 28.4  
4:00 272.1 F 29.2

Min 271.1 F 28.4  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.9 F 0.0  
5:00 231.4 F 0.0  
10:00 224.2 F 0.0  
15:00 224.6 F 0.0  
20:00 224.6 F 0.0  
25:00 220.6 F 0.0  
30:00 217.8 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 23:08  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:02:06  
CYCLE COMPLETE

03/27/2024 01:16 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_