
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005528
10:32 15/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:04
PRESSURIZING 2:08
STERILIZING 3:08
136.5 C 224kPa 3:08
136.1 C 222kPa 3:38
136.4 C 221kPa 4:08
136.4 C 224kPa 4:38
136.4 C 225kPa 5:08
136.1 C 219kPa 5:38
136.4 C 220kPa 6:08
136.1 C 223kPa 6:38
136.1 C 223kPa 6:38
VENTING 6:38
AIR DRYING 7:31
CYCLE COMPLETE 11:11

Digital Signature #
DCF2425A3E428E8A
