
Midmark M11 - v1.0.5
Total Cycles: 4157

Name: RITTER2
Sterilizer ID: V2313228
11/26/2024 10:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.1 F 0.0
2:00 78.8 F 0.0
4:00 76.7 F 0.0
6:00 99.4 F 0.1
8:00 144.0 F 0.2
10:00 183.5 F 0.4
12:00 201.7 F 1.8
14:00 211.4 F 4.6
16:00 220.3 F 7.7
18:00 227.2 F 11.2
20:00 230.6 F 7.1
22:00 239.2 F 11.7
24:00 236.9 F 9.7
26:00 239.6 F 9.9
28:00 248.6 F 14.9
30:00 259.9 F 21.2
32:00 270.2 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.1 F 29.0
2:00 272.5 F 29.5
3:00 271.8 F 29.2
4:00 271.9 F 29.0

Min 271.5 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.9 F 0.0
5:00 215.0 F 0.0
10:00 214.7 F 0.0
15:00 222.0 F 0.0
20:00 228.3 F 0.0
25:00 239.2 F 0.0
30:00 228.0 F 0.0

Drying Complete

FILLING: 0:56
HEATING: 32:50
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:11:35
CYCLE COMPLETE

11/26/2024 11:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____