
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001077
16:05 13/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:44
STERILIZING 3:55
136.5 C 224kPa 3:55
136.4 C 225kPa 4:25
136.2 C 223kPa 4:55
136.3 C 224kPa 5:25
136.4 C 225kPa 5:55
136.3 C 224kPa 6:25
136.2 C 222kPa 6:55
136.4 C 225kPa 7:25
136.4 C 225kPa 7:25
VENTING 7:25
AIR DRYING 8:16
CYCLE COMPLETE 23:16

Digital Signature #
764ED00D21AFF54D
