
Midmark M11 - v1.0.5
Total Cycles: 498

Name: RITTER2
Sterilizer ID: V2313228
10/28/2021 12:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.2 F 0.0
2:00 159.4 F 2.9
4:00 176.0 F 6.9
6:00 196.6 F 12.1
8:00 218.0 F 6.2
10:00 231.2 F 11.4
12:00 233.8 F 8.9
14:00 239.6 F 10.3
16:00 247.8 F 14.5
18:00 260.5 F 21.4
20:00 271.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.4 F 29.4
2:00 271.6 F 29.1
3:00 272.9 F 29.6
4:00 271.7 F 28.9

Min 271.5 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.7 F 0.0
5:00 238.2 F 0.0
10:00 220.0 F 0.0
15:00 221.6 F 0.0
20:00 225.2 F 0.0
25:00 228.5 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 20:37
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:59:20
CYCLE COMPLETE

10/28/2021 01:13 PM
ACCEPTED: X
REJECTED:
Operator: U02