
Midmark M11 - v1.0.0
Total Cycles: 5423

Name: RITTER 1
Sterilizer ID: V2178198
03/19/2025 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.0 F 0.0
2:00 96.1 F 1.0
4:00 124.6 F 3.4
6:00 155.7 F 6.5
8:00 178.9 F 10.0
10:00 201.4 F 5.8
12:00 214.7 F 5.4
14:00 226.2 F 9.5
16:00 235.3 F 11.0
18:00 237.5 F 10.1
20:00 236.0 F 8.4
22:00 247.4 F 13.9
24:00 259.0 F 20.3
26:00 269.8 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.5 F 28.8
2:00 272.4 F 29.0
3:00 272.1 F 28.5
4:00 272.7 F 29.2

Min 271.6 F 28.3
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.1 F 0.0
5:00 179.9 F 0.0
10:00 191.9 F 0.0
15:00 196.3 F 0.0
20:00 196.8 F 0.0
25:00 191.8 F 0.0
30:00 189.8 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 26:56
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:05:13
CYCLE COMPLETE

03/19/2025 01:30 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____