
Midmark M11 - v1.0.5
Total Cycles: 4042

Name: RITTER2
Sterilizer ID: V2313228
10/18/2024 11:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 87.8 F 0.0
2:00 89.5 F 0.0
4:00 104.9 F 0.0
6:00 142.9 F 0.2
8:00 172.2 F 2.4
10:00 190.0 F 5.4
12:00 203.5 F 8.5
14:00 214.5 F 12.5
16:00 221.0 F 6.3
18:00 231.9 F 10.5
20:00 233.2 F 8.4
22:00 243.2 F 12.8
24:00 242.2 F 11.5
26:00 254.0 F 17.5
28:00 264.6 F 23.9
30:00 271.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.8 F 29.4
2:00 271.9 F 28.6
3:00 272.1 F 29.2
4:00 272.7 F 29.2

Min 271.5 F 28.5
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.5 F 0.0
5:00 163.7 F 0.0
10:00 185.0 F 0.0
15:00 186.3 F 0.0
20:00 183.9 F 0.0
25:00 183.9 F 0.0
30:00 193.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:04
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:08:34
CYCLE COMPLETE

10/18/2024 12:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____