
Midmark M11 - v1.0.0
Total Cycles: 2911

Name: RITTER 1
Sterilizer ID: V2178198
02/16/2023 09:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.8 F 0.0
2:00 79.5 F 0.0
4:00 88.4 F 0.6
6:00 120.5 F 2.9
8:00 151.9 F 6.0
10:00 171.6 F 9.3
12:00 198.3 F 13.0
14:00 207.3 F 4.8
16:00 220.8 F 8.3
18:00 229.0 F 12.2
20:00 231.7 F 7.7
22:00 241.8 F 12.8
24:00 239.5 F 10.4
26:00 251.5 F 16.3
28:00 262.2 F 22.4
30:00 272.1 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.6 F 29.3
2:00 271.8 F 28.6
3:00 272.0 F 29.2
4:00 272.4 F 29.0

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 221.2 F 0.0
10:00 219.5 F 0.0
15:00 216.3 F 0.0
20:00 227.8 F 0.0
25:00 236.4 F 0.0
30:00 232.6 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 30:32
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:08:31
CYCLE COMPLETE

02/16/2023 10:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____