
Midmark M11 - v1.0.0
Total Cycles: 1435

Name: RITTER 1
Sterilizer ID: V2178198
12/08/2021 02:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.5 F 0.0
2:00 182.9 F 2.8
4:00 193.6 F 7.2
6:00 217.5 F 12.8
8:00 229.3 F 7.5
10:00 241.2 F 13.0
12:00 240.4 F 10.6
14:00 236.9 F 9.0
16:00 252.1 F 16.1
18:00 263.9 F 23.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.7 F 29.0
2:00 272.6 F 29.1
3:00 272.1 F 28.9
4:00 271.7 F 28.4

Min 270.8 F 27.9
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 231.3 F 0.0
10:00 246.7 F 0.0
15:00 242.9 F 0.0
20:00 224.5 F 0.0
25:00 228.8 F 0.0
30:00 227.5 F 0.0

Drying Complete

FILLING: 1:38
HEATING: 19:54
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

12/08/2021 03:59 PM
ACCEPTED: X
REJECTED:
Operator: U02