
Midmark M11 - v1.0.0
Total Cycles: 2569

Name: RITTER 1
Sterilizer ID: V2178198
11/16/2022 02:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.2 F 0.0
2:00 199.1 F 4.8
4:00 215.8 F 9.9
6:00 220.5 F 5.8
8:00 232.2 F 10.7
10:00 231.6 F 7.7
12:00 242.1 F 11.4
14:00 245.8 F 13.1
16:00 258.5 F 20.1
18:00 270.0 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.4
1:00 272.2 F 29.2
2:00 272.0 F 28.6
3:00 271.7 F 28.9
4:00 272.3 F 29.0

Min 271.3 F 28.3
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 225.5 F 0.0
10:00 233.8 F 0.0
15:00 229.7 F 0.0
20:00 230.6 F 0.0
25:00 229.6 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 18:45
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:41
CYCLE COMPLETE

11/16/2022 03:52 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____