
Midmark M11 - v1.0.5
Total Cycles: 1442

Name: RITTER2
Sterilizer ID: V2313228
08/31/2022 03:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.8 F 0.0
2:00 181.2 F 2.0
4:00 192.7 F 5.3
6:00 220.7 F 10.0
8:00 219.7 F 6.9
10:00 228.0 F 10.1
12:00 231.2 F 8.3
14:00 241.6 F 12.9
16:00 241.3 F 11.7
18:00 253.8 F 18.1
20:00 265.3 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.5 F 29.5
2:00 272.3 F 30.0
3:00 271.8 F 29.2
4:00 272.7 F 29.9

Min 271.6 F 29.2
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.7 F 0.0
5:00 217.1 F 0.0
10:00 231.4 F 0.0
15:00 223.9 F 0.0
20:00 228.5 F 0.0
25:00 217.1 F 0.0
30:00 215.6 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 21:48
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:14
CYCLE COMPLETE

08/31/2022 04:31 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____