
Midmark M11 - v1.0.0
Total Cycles: 3674

Name: RITTER 1
Sterilizer ID: V2178198
08/18/2023 08:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.3 F 0.0
2:00 78.0 F 0.0
4:00 99.8 F 0.0
6:00 125.2 F 0.6
8:00 153.0 F 3.0
10:00 173.3 F 5.9
12:00 191.8 F 9.0
14:00 205.1 F 12.2
16:00 213.6 F 4.3
18:00 222.5 F 7.7
20:00 233.5 F 11.4
22:00 230.5 F 7.1
24:00 242.1 F 12.4
26:00 239.4 F 10.3
28:00 251.4 F 16.1
30:00 261.7 F 22.0
32:00 271.3 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 28.9
2:00 271.8 F 28.7
3:00 272.2 F 28.7
4:00 272.0 F 28.9

Min 271.4 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.8 F 0.0
5:00 204.8 F 0.0
10:00 214.5 F 0.0
15:00 224.8 F 0.0
20:00 226.3 F 0.0
25:00 230.1 F 0.0
30:00 228.3 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:39
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:10:56
CYCLE COMPLETE

08/18/2023 09:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____