
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000470
8:09 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:49
STERILIZING 3:54
136.0 C 220kPa 3:54
136.3 C 221kPa 4:24
136.4 C 224kPa 4:54
136.4 C 222kPa 5:24
136.1 C 223kPa 5:54
CYCLE ABORTED 6:08
STOP BUTTON PRESSED
NOT STERILE
END TIME 6:08

Digital Signature #
512D866D96538076
