

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2137

Name: RITTER2  
Sterilizer ID: V2313228  
03/09/2023 03:56 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.2 F 0.0  
2:00 182.3 F 4.0  
4:00 204.3 F 9.5  
6:00 220.2 F 6.8  
8:00 232.7 F 11.4  
10:00 236.2 F 9.6  
12:00 237.1 F 9.0  
14:00 251.5 F 16.3  
16:00 264.1 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.7  
1:00 272.6 F 29.6  
2:00 272.2 F 29.1  
3:00 271.2 F 28.6  
4:00 271.8 F 29.5  
  
Min 271.1 F 28.6  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.0 F 0.0  
5:00 230.9 F 0.0  
10:00 225.4 F 0.0  
15:00 222.0 F 0.0  
20:00 217.7 F 0.0  
25:00 222.7 F 0.0  
30:00 224.0 F 0.0  
Drying Complete

FILLING: 2:17  
HEATING: 17:58  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:57:31  
CYCLE COMPLETE

03/09/2023 04:53 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_