
Midmark M11 - v1.0.0
Total Cycles: 5319

Name: RITTER 1
Sterilizer ID: V2178198
02/06/2025 10:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.7 F 0.0
2:00 80.3 F 0.5
4:00 95.3 F 2.0
6:00 145.1 F 4.2
8:00 153.4 F 7.3
10:00 176.4 F 11.0
12:00 199.7 F 2.7
14:00 215.2 F 4.8
16:00 224.2 F 8.1
18:00 234.1 F 11.9
20:00 231.7 F 6.9
22:00 243.6 F 12.1
24:00 238.6 F 9.6
26:00 251.5 F 15.8
28:00 262.5 F 22.0
30:00 272.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.5
1:00 272.1 F 28.4
2:00 272.4 F 28.5
3:00 272.7 F 28.8
4:00 272.7 F 29.1

Min 271.9 F 28.4
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.9 F 0.0
5:00 211.1 F 0.0
10:00 223.7 F 0.0
15:00 227.9 F 0.0
20:00 228.5 F 0.0
25:00 229.5 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 31:05
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:09:22
CYCLE COMPLETE

02/06/2025 11:37 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____