
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001197
11:56 28/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:37
STERILIZING 5:13
136.4 C 225kPa 5:13
136.2 C 224kPa 5:43
136.3 C 224kPa 6:13
136.3 C 224kPa 6:43
136.3 C 224kPa 7:13
136.2 C 224kPa 7:43
136.3 C 225kPa 8:13
136.4 C 224kPa 8:43
136.4 C 224kPa 8:43
VENTING 8:43
AIR DRYING 9:41
CYCLE COMPLETE 11:06

Digital Signature #
38C3930F3CDF6945
