
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001963
9:51 16/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:30
PRESSURIZING 1:36
STERILIZING 2:40
136.4 C 224kPa 2:40
136.4 C 227kPa 3:10
136.3 C 224kPa 3:40
136.3 C 225kPa 4:10
136.1 C 224kPa 4:40
136.2 C 225kPa 5:10
136.4 C 227kPa 5:40
136.3 C 225kPa 6:10
136.3 C 225kPa 6:10
VENTING 6:10
AIR DRYING 6:59
CYCLE COMPLETE 36:47

Digital Signature #
F7A5040E73E78A12
