
Midmark M11 - v1.0.0
Total Cycles: 2313

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2022 03:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.6 F 0.0
2:00 179.2 F 4.1
4:00 208.3 F 8.5
6:00 216.9 F 13.3
8:00 222.9 F 7.5
10:00 236.1 F 12.7
12:00 236.4 F 9.9
14:00 237.0 F 8.4
16:00 248.5 F 14.8
18:00 260.5 F 21.6
20:00 271.8 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.3 F 29.1
2:00 271.8 F 29.0
3:00 272.2 F 29.1
4:00 271.6 F 28.9

Min 271.4 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 232.3 F 0.0
10:00 234.3 F 0.0
15:00 230.3 F 0.0
20:00 228.4 F 0.0
25:00 230.9 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 20:33
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:59
CYCLE COMPLETE

09/12/2022 04:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____