
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000813
16:35 23/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 4:19
STERILIZING 5:42
136.0 C 220kPa 5:42
136.4 C 219kPa 6:12
135.8 C 219kPa 6:42
136.3 C 222kPa 7:12
136.2 C 218kPa 7:42
136.4 C 223kPa 8:12
136.4 C 222kPa 8:42
136.2 C 220kPa 9:12
136.2 C 220kPa 9:12
VENTING 9:12
AIR DRYING 10:20
CYCLE COMPLETE 40:21

Digital Signature #
EB552F0A545F9A2B
