
Midmark M11 - v1.0.0
Total Cycles: 3565

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2023 06:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.9 F 0.0
2:00 182.2 F 1.5
4:00 198.5 F 5.6
6:00 216.3 F 11.1
8:00 225.3 F 6.8
10:00 238.0 F 12.3
12:00 237.7 F 10.0
14:00 235.6 F 8.2
16:00 250.5 F 15.5
18:00 262.9 F 22.5
20:00 271.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.2 F 28.5
2:00 272.4 F 29.0
3:00 271.6 F 28.1
4:00 272.6 F 28.9

Min 271.3 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.9 F 0.0
5:00 211.6 F 0.0
10:00 218.4 F 0.0
15:00 224.5 F 0.0
20:00 224.5 F 0.0
25:00 224.2 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 20:07
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:06
CYCLE COMPLETE

07/21/2023 07:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____