

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2403

Name: RITTER2  
Sterilizer ID: V2313228  
05/23/2023 04:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.4 F 0.0  
2:00 176.6 F 3.8  
4:00 189.0 F 8.4  
6:00 209.8 F 12.8  
8:00 224.3 F 8.3  
10:00 237.1 F 12.4  
12:00 239.5 F 11.5  
14:00 238.7 F 10.2  
16:00 252.2 F 16.9  
18:00 264.2 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.6 F 29.6  
2:00 271.5 F 28.9  
3:00 272.7 F 29.8  
4:00 272.1 F 29.1  
  
Min 271.5 F 28.8  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.6 F 0.0  
5:00 212.2 F 0.0  
10:00 224.1 F 0.0  
15:00 228.2 F 0.0  
20:00 227.7 F 0.0  
25:00 225.0 F 0.0  
30:00 223.0 F 0.0  
Drying Complete

FILLING: 2:02  
HEATING: 19:56  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:11  
CYCLE COMPLETE

05/23/2023 05:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_