
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004564
14:42 18/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:01
STERILIZING 2:56
136.4 C 220kPa 2:56
136.4 C 224kPa 3:26
136.1 C 222kPa 3:56
136.2 C 220kPa 4:26
136.1 C 223kPa 4:56
136.2 C 220kPa 5:26
136.3 C 223kPa 5:56
136.4 C 224kPa 6:26
136.4 C 224kPa 6:26
VENTING 6:26
AIR DRYING 7:17
CYCLE COMPLETE 7:19

Digital Signature #
E66D80FBE6E1C545
