
Midmark M11 - v1.0.0
Total Cycles: 3197

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2023 02:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 214.3 F 0.0
2:00 211.1 F 4.2
4:00 220.1 F 9.8
6:00 231.8 F 6.7
8:00 237.7 F 11.8
10:00 237.0 F 9.7
12:00 237.9 F 8.6
14:00 250.7 F 15.7
16:00 263.1 F 22.9
18:00 271.7 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 29.0
2:00 271.5 F 28.4
3:00 272.5 F 29.1
4:00 271.7 F 28.4

Min 271.4 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.2 F 0.0
5:00 218.3 F 0.0
10:00 232.7 F 0.0
15:00 234.7 F 0.0
20:00 228.8 F 0.0
25:00 222.3 F 0.0
30:00 217.5 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 18:05
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:57:39
CYCLE COMPLETE

05/03/2023 03:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____